



Full Length Article



The JUNO experiment Top Tracker

The JUNO Collaboration, Angel Abusleme^{6,5}, Thomas Adam⁴⁵, Shakeel Ahmad⁶⁹, Rizwan Ahmed⁶⁹, Sebastiano Aiello⁵⁸, Muhammad Akram⁶⁹, Abid Aleem⁶⁹, Tsagkarakis Alexandros⁵¹, Fengpeng An²¹, Qi An²³, Giuseppe Andronico⁵⁸, Nikolay Anfimov⁷⁰, Vito Antonelli⁶⁰, Tatiana Antoshkina⁷⁰, Burin Asavapibhop⁷⁴, João Pedro Athayde Marcondes de André^{45,*}, Didier Auguste⁴³, Weidong Bai²¹, Nikita Balashov⁷⁰, Wander Baldini⁵⁹, Andrea Barresi⁶¹, Davide Basilico⁶⁰, Eric Baussan⁴⁵, Marco Bellato⁶³, Marco Beretta⁶⁰, Antonio Bergnoli⁶³, Daniel Bick⁵², Thilo Birkenfeld⁵¹, Sylvie Blin⁴⁸, David Blum⁵⁷, Simon Blyth¹¹, Anastasia Bolshakova⁷⁰, Mathieu Bongrand⁴⁷, Clément Bordereau^{44,40}, Dominique Breton⁴³, Augusto Brigatti⁶⁰, Riccardo Brugnera^{64,63}, Riccardo Bruno⁵⁸, Antonio Budano⁶⁷, Jose Busto⁴⁶, Anatael Cabrera⁴³, Barbara Caccianiga⁶⁰, Hao Cai³⁴, Xiao Cai¹¹, Yanke Cai¹¹, Zhiyan Cai¹¹, Stéphane Callier⁴⁴, Antonio Cammi⁶², Agustin Campeny^{6,5}, Chuanya Cao¹¹, Guofu Cao¹¹, Jun Cao¹¹, Rossella Caruso⁵⁸, Cédric Cerna⁴⁴, Vanessa Cerrone^{64,63}, Chi Chan³⁸, Jinfan Chang¹¹, Yun Chang³⁹, Chao Chen¹¹, Guoming Chen²⁸, Pingping Chen¹⁹, Shaomin Chen¹⁴, Yixue Chen¹², Yu Chen²¹, Zhiyuan Chen¹¹, Zikang Chen²¹, Jie Cheng¹², Yaping Cheng⁸, Yu Chin Cheng⁴⁰, Alexander Chepurinov⁷², Alexey Chetverikov⁷⁰, Davide Chiesa⁶¹, Pietro Chimenti³, Ziliang Chu¹¹, Artem Chukanov⁷⁰, Gérard Claverie⁴⁴, Catia Clementi⁶⁵, Barbara Clerbaux², Marta Colomer Molla², Selma Conforti Di Lorenzo⁵⁰, Alberto Coppi^{64,63}, Daniele Corti⁶³, Flavio Dal Corso⁶³, Olivia Dalager⁷⁷, Christophe De La Taille⁵⁰, Zhi Deng¹⁴, Ziyan Deng¹¹, Wilfried Depnering⁵⁴, Marco Diaz⁶, Xuefeng Ding¹¹, Yayun Ding¹¹, Bayu Dirgantara⁷⁶, Sergey Dmitrievsky⁷⁰, Tadeas Dohnal⁴¹, Dmitry Dolzhikov⁷⁰, Georgy Donchenko⁷², Jianmeng Dong¹⁴, Evgeny Doroshkevich⁷¹, Wei Dou¹⁴, Marcos Dracos^{45,*}, Olivier Drapier⁴⁹, Frédéric Druillolle⁴⁴, Ran Du¹¹, Shuxian Du³⁷, Katherine Dugas⁷⁷, Stefano Dusini⁶³, Hongyue Duyang²⁶, Jessica Eck⁵⁷, Timo Enqvist⁴², Andrea Fabbri⁶⁷, Ulrike Fahrenholz⁵⁵, Lei Fan¹¹, Jian Fang¹¹, Wenxing Fang¹¹, Marco Fargetta⁵⁸, Dmitry Fedoseev⁷⁰, Zhengyong Fei¹¹, Giulietto Felici⁶⁶, Li-Cheng Feng³⁸, Qichun Feng²², Federico Ferraro⁶⁰, Amélie Fournier⁴⁴, Haonan Gan³², Feng Gao⁵¹, Alberto Garfagnini^{64,63}, Arsenii Gavrikov^{64,63}, Vladimir Gerasimov⁷⁰, Marco Giammarchi⁶⁰, Nunzio Giudice⁵⁸, Maxim Gonchar⁷⁰, Guanghua Gong¹⁴, Hui Gong¹⁴, Yuri Gornushkin⁷⁰, Alexandre Göttel^{53,51}, Marco Grassi^{64,63}, Maxim Gromov⁷², Vasily Gromov⁷⁰, Minghao Gu¹¹, Xiaofei Gu³⁷, Yu Gu²⁰, Mengyun Guan¹¹, Yuduo Guan¹¹, Nunzio Guardone⁵⁸, Cong Guo¹¹, Wanlei Guo¹¹, Xinheng Guo⁹, Yuhang Guo³⁵, Semen Gursky⁷⁰, Caren Hagner⁵², Ran Han⁸, Yang Han²¹, Miao He¹¹, Wei He¹¹, Tobias Heinz⁵⁷, Patrick Hellmuth⁴⁴, Yuekun Heng¹¹, Rafael Herrera^{6,5}, YuenKeung Hor²¹, Shaojing Hou¹¹, Yee Hsiung⁴⁰, Bei-Zhen Hu⁴⁰, Hang Hu²¹, Jianrun Hu¹¹, Jun Hu¹¹,

* Corresponding author.

** Now at Istituto Superiore per la Protezione e la Ricerca Ambientale, 00144 Rome, Italy

E-mail addresses: juno_pub_comm@juno.ihep.ac.cn (The JUNO Collaboration), jpandre@iphc.cnrs.fr (J.P.A.M. de André), marcos.dracos@in2p3.fr (M. Dracos).

<https://doi.org/10.1016/j.nima.2023.168680>

Received 20 March 2023; Received in revised form 28 July 2023; Accepted 11 September 2023

Available online 15 September 2023

0168-9002/© 2023 Elsevier B.V. All rights reserved.

Shouyang Hu¹⁰, Tao Hu¹¹, Yuxiang Hu¹¹, Zhuojun Hu²¹, Guihong Huang²⁵, Hanxiong Huang¹⁰, Jinhao Huang¹¹, Kaixuan Huang²¹, Wenhao Huang²⁶, Qinhuang Huang^{49,45}, Xin Huang¹¹, Xingtao Huang²⁶, Yongbo Huang²⁸, Jiaqi Hui³⁰, Lei Huo²², Wenju Huo²³, Cédric Huss⁴⁴, Safeer Hussain⁶⁹, Ara Ioannisian¹, Roberto Isocrate⁶³, Beatrice Jelmini^{64,63}, Ignacio Jeria⁶, Xiaolu Ji¹¹, Huihui Jia³³, Junji Jia³⁴, Siyu Jian¹⁰, Di Jiang²³, Wei Jiang¹¹, Xiaoshan Jiang¹¹, Xiaoping Jing¹¹, Cécile Jollet⁴⁴, Leonidas Kalousis⁴⁵, Philipp Kampmann^{56,53}, Li Kang¹⁹, Rebin Karaparambil⁴⁵, Narine Kazarian¹, Ali Khan⁶⁹, Amina Khatun⁷³, Khanchai Khosonthongkee⁷⁶, Denis Korablev⁷⁰, Konstantin Kouzakov⁷², Alexey Krasnoperov⁷⁰, Sergey Kuleshov⁵, Nikolay Kutovskiy⁷⁰, Pasi Kuusiniemi⁴², Tobias Lachenmaier⁵⁷, Cecilia Landini⁶⁰, Sébastien Leblanc⁴⁴, Victor Lebrin⁴⁷, Frederic Lefevre⁴⁷, Ruiting Lei¹⁹, Rupert Leitner⁴¹, Jason Leung³⁸, Demin Li³⁷, Fei Li¹¹, Fule Li¹⁴, Gaosong Li¹¹, Huiling Li¹¹, Mengzhao Li¹¹, Min Li¹¹, Nan Li¹⁷, Qingjiang Li¹⁷, Ruhui Li¹¹, Rui Li³⁰, Shanfeng Li¹⁹, Tao Li²¹, Teng Li²⁶, Weidong Li^{11,15}, Weiguo Li¹¹, Xiaomei Li¹⁰, Xiaonan Li¹¹, Xinglong Li¹⁰, Yi Li¹⁹, Yichen Li¹¹, Yufeng Li¹¹, Zepeng Li¹¹, Zhaohan Li¹¹, Zhibing Li²¹, Ziyuan Li²¹, Zonghai Li³⁴, Hao Liang¹⁰, Hao Liang²³, Jiajun Liao²¹, Ayut Limphirat⁷⁶, Guey-Lin Lin³⁸, Shengxin Lin¹⁹, Tao Lin¹¹, Jiajie Ling²¹, Ivano Lippi⁶³, Caimei Liu¹¹, Fang Liu¹², Haidong Liu³⁷, Haotian Liu³⁴, Hongbang Liu²⁸, Hongjuan Liu²⁴, Hongtao Liu²¹, Hui Liu²⁰, Jianglai Liu^{30,31}, Jiayi Liu¹¹, Jinchang Liu¹¹, Min Liu²⁴, Qian Liu¹⁵, Qin Liu²³, Runxuan Liu^{53,51}, Shenghui Liu¹¹, Shubin Liu²³, Shulin Liu¹¹, Xiaowei Liu²¹, Xiwen Liu²⁸, Yankai Liu³⁵, Yunzhe Liu¹¹, Alexey Lokhov^{72,71}, Paolo Lombardi⁶⁰, Claudio Lombardo⁵⁸, Kai Loo⁴², Chuan Lu³², Haoqi Lu¹¹, Jingbin Lu¹⁶, Junguang Lu¹¹, Peizhi Lu²¹, Shuxiang Lu³⁷, Bayarto Lubsandorzhev⁷¹, Sultim Lubsandorzhev⁷¹, Livia Ludhova^{53,51}, Arslan Lukanov⁷¹, Daibin Luo¹¹, Fengjiao Luo²⁴, Guang Luo²¹, Jianyi Luo²¹, Shu Luo³⁶, Wuming Luo¹¹, Xiaojie Luo¹¹, Vladimir Lyashuk⁷¹, Bangzheng Ma²⁶, Bing Ma³⁷, Qiumei Ma¹¹, Si Ma¹¹, Xiaoyan Ma¹¹, Xubo Ma¹², Jihane Maalmi⁴³, Marco Magoni⁶⁰, Jingyu Mai²¹, Yury Malyshev^{56,53}, Roberto Carlos Mandujano⁷⁷, Fabio Mantovani⁵⁹, Xin Mao⁸, Yajun Mao¹³, Stefano M. Mari⁶⁷, Filippo Marini^{64,63}, Agnese Martini⁶⁶, Matthias Mayer⁵⁵, Davit Mayilyan¹, Ints Mednieks⁶⁸, Yue Meng³⁰, Anita Meraviglia^{56,53,51}, Anselmo Mereaglia⁴⁴, Emanuela Meroni⁶⁰, David Meyhöfer⁵², Jonathan Miller⁷, Lino Miramonti⁶⁰, Paolo Montini⁶⁷, Michele Montuschi⁵⁹, Axel Müller⁵⁷, Massimiliano Nastasi⁶¹, Dmitry V. Naumov⁷⁰, Elena Naumova⁷⁰, Diana Navas-Nicolas⁴³, Igor Nemchenok⁷⁰, Minh Thuan Nguyen Thi³⁸, Alexey Nikolaev⁷², Feipeng Ning¹¹, Zhe Ning¹¹, Hiroshi Nunokawa⁴, Lothar Oberauer⁵⁵, Juan Pedro Ochoa-Ricoux^{77,6,5}, Alexander Olshevskiy⁷⁰, Domizia Orestano⁶⁷, Fausto Ortica⁶⁵, Rainer Othegraven⁵⁴, Alessandro Paoloni⁶⁶, Sergio Parmeggiano⁶⁰, Yatian Pei¹¹, Luca Pelicci^{53,51}, Anguo Peng²⁴, Haiping Peng²³, Yu Peng¹¹, Zhaoyuan Peng¹¹, Frédéric Perrot⁴⁴, Pierre-Alexandre Petitjean², Fabrizio Petrucci⁶⁷, Oliver Pilarczyk⁵⁴, Luis Felipe Piñeres Rico⁴⁵, Artyom Popov⁷², Pascal Poussot⁴⁵, Ezio Previtali⁶¹, Fazhi Qi¹¹, Ming Qi²⁷, Sen Qian¹¹, Xiaohui Qian¹¹, Zhen Qian²¹, Hao Qiao¹³, Zhonghua Qin¹¹, Shoukang Qiu²⁴, Gioacchino Ranucci⁶⁰, Reem Rasheed⁴⁴, Alessandra Re⁶⁰, Abdel Rebi⁴⁴, Mariia Redchuk⁶³, Bin Ren¹⁹, Jie Ren¹⁰, Barbara Ricci⁵⁹, Mariam Rifai^{53,51}, Mathieu Roche⁴⁴, Narongkiat Rodphai¹¹, Aldo Romani⁶⁵, Victor Romanov⁷⁰, Bedřich Roskovec⁴¹, Xichao Ruan¹⁰, Arseniy Rybnikov⁷⁰, Andrey Sadovsky⁷⁰, Paolo Saggese⁶⁰, Deshan Sandanayake⁴⁵, Simone Sanfilippo⁶⁷, Anut Sangka⁷⁵, Utane Sawangwit⁷⁵, Julia Sawatzki⁵⁵, Michaela Schever^{53,51}, Jacky Schuler⁴⁵, Cédric Schwab⁴⁵, Konstantin Schweizer⁵⁵, Alexandr Selyunin⁷⁰, Andrea Serafini^{64,63}, Giulio Settanta^{53,**}, Mariangela Settimo⁴⁷, Vladislav Sharov⁷⁰, Arina Shaydurova⁷⁰, Jingyan Shi¹¹, Yanan Shi¹¹, Vitaly Shutov⁷⁰, Andrey Sidorenkov⁷¹, Fedor Šimkovic⁷³, Chiara Sirignano^{64,63}, Jaruchit Siripak⁷⁶, Monica Sisti⁶¹, Maciej Slupecki⁴², Mikhail Smirnov²¹, Oleg Smirnov⁷⁰, Thiago Sogo-Bezerra⁴⁷, Sergey Sokolov⁷⁰, Julianan Songwadhana⁷⁶, Boonrucksar Soonthornthum⁷⁵, Albert Sotnikov⁷⁰, Ondřej Šrámek⁴¹, Warintorn Sreethawong⁷⁶, Achim Stahl⁵¹, Luca Stanco⁶³, Konstantin Stankevich⁷², Dušan Štefánik⁷³, Hans Steiger^{54,55}, Jochen Steinmann⁵¹, Tobias Sterr⁵⁷, Matthias Raphael Stock⁵⁵, Virginia Strati⁵⁹, Alexander Studenikin⁷², Jun Su²¹,

Shifeng Sun¹², Xilei Sun¹¹, Yongjie Sun²³, Yongzhao Sun¹¹, Zhengyang Sun³¹,
 Narumon Suwonjandee⁷⁴, Michal Szelezniak⁴⁵, Akira Takenaka³¹, Jian Tang²¹, Qiang Tang²¹,
 Quan Tang²⁴, Xiao Tang¹¹, Vidhya Thara Hariharan⁵², Eric Theisen⁵⁴, Alexander Tietzsch⁵⁷,
 Igor Tkachev⁷¹, Tomas Tmej⁴¹, Marco Danilo Claudio Torri⁶⁰, Francesco Tortorici⁵⁸,
 Konstantin Treskov⁷⁰, Andrea Triossi^{64,63}, Riccardo Triozzi^{64,63}, Giancarlo Troni⁶,
 Wladyslaw Trzaska⁴², Yu-Chen Tung⁴⁰, Cristina Tuve⁵⁸, Nikita Ushakov⁷¹, Vadim Vedin⁶⁸,
 Giuseppe Verde⁵⁸, Maxim Vialkov⁷², Benoit Viaud⁴⁷, Cornelius Moritz Vollbrecht^{53,51},
 Katharina von Sturm^{64,63}, Vit Vorobel⁴¹, Dmitriy Voronin⁷¹, Lucia Votano⁶⁶, Pablo Walker^{6,5},
 Caishen Wang¹⁹, Chung-Hsiang Wang³⁹, En Wang³⁷, Guoli Wang²², Jian Wang²³, Jun Wang²¹,
 Lu Wang¹¹, Meng Wang²⁴, Meng Wang²⁶, Ruiguang Wang¹¹, Siguang Wang¹³, Wei Wang²¹,
 Wenshuai Wang¹¹, Xi Wang¹⁷, Xiangyue Wang²¹, Yangfu Wang¹¹, Yaoguang Wang¹¹,
 Yi Wang¹¹, Yi Wang¹⁴, Yifang Wang¹¹, Yuanqing Wang¹⁴, Zhe Wang¹⁴, Zheng Wang¹¹,
 Zhimin Wang¹¹, Apimook Watcharangkool⁷⁵, Wei Wei¹¹, Wei Wei²⁶, Wenlu Wei¹¹,
 Yadong Wei¹⁹, Kaile Wen¹¹, Liangjian Wen¹¹, Jun Weng¹⁴, Christopher Wiebusch⁵¹,
 Rosmarie Wirth⁵², Bjoern Wonsak⁵², Diru Wu¹¹, Qun Wu²⁶, Zhi Wu¹¹, Michael Wurm⁵⁴,
 Jacques Wurtz⁴⁵, Christian Wysotzki⁵¹, Yufei Xi³², Dongmei Xia¹⁸, Xiang Xiao²¹,
 Xiaochuan Xie²⁸, Yuguang Xie¹¹, Zhangquan Xie¹¹, Zhao Xin¹¹, Zhizhong Xing¹¹, Benda Xu¹⁴,
 Cheng Xu²⁴, Donglian Xu^{31,30}, Fanrong Xu²⁰, Hangkun Xu¹¹, Jilei Xu¹¹, Jing Xu⁹,
 Meihang Xu¹¹, Yin Xu³³, Yu Xu²¹, Baojun Yan¹¹, Qiyu Yan¹⁵, Taylor Yan⁷⁶, Xiongbo Yan¹¹,
 Yupeng Yan⁷⁶, Changgen Yang¹¹, Chengfeng Yang²⁸, Jie Yang³⁷, Lei Yang¹⁹, Xiaoyu Yang¹¹,
 Yifan Yang¹¹, Yifan Yang², Haifeng Yao¹¹, Jiaxuan Ye¹¹, Mei Ye¹¹, Ziping Ye³¹,
 Frédéric Yermia⁴⁷, Zhengyun You²¹, Boxiang Yu¹¹, Chiye Yu¹⁹, Chunxu Yu³³, Guojun Yu²⁷,
 Hongzhao Yu²¹, Miao Yu³⁴, Xianghui Yu³³, Zeyuan Yu¹¹, Zezhong Yu¹¹, Cenxi Yuan²¹,
 Chengzhuo Yuan¹¹, Ying Yuan¹³, Zhenxiong Yuan¹⁴, Baobiao Yue²¹, Noman Zafar⁶⁹,
 Vitalii Zavadskiy⁷⁰, Shan Zeng¹¹, Tingxuan Zeng¹¹, Yuda Zeng²¹, Liang Zhan¹¹,
 Aiqiang Zhang¹⁴, Bin Zhang³⁷, Binting Zhang¹¹, Feiyang Zhang³⁰, Haosen Zhang¹¹,
 Honghao Zhang²¹, Jialiang Zhang²⁷, Jiawen Zhang¹¹, Jie Zhang¹¹, Jin Zhang²⁸,
 Jingbo Zhang²², Jinnan Zhang¹¹, Mohan Zhang¹¹, Peng Zhang¹¹, Qingmin Zhang³⁵,
 Shiqi Zhang²¹, Shu Zhang²¹, Shuihan Zhang¹¹, Tao Zhang³⁰, Xiaomei Zhang¹¹, Xin Zhang¹¹,
 Xuantong Zhang¹¹, Yinhong Zhang¹¹, Yiyu Zhang¹¹, Yongpeng Zhang¹¹, Yu Zhang¹¹,
 Yuanyuan Zhang³¹, Yumei Zhang²¹, Zhenyu Zhang³⁴, Zhijian Zhang¹⁹, Jie Zhao¹¹,
 Rong Zhao²¹, Runze Zhao¹¹, Shujun Zhao³⁷, Dongqin Zheng²⁰, Hua Zheng¹⁹,
 Yangheng Zheng¹⁵, Weirong Zhong²⁰, Jing Zhou¹⁰, Li Zhou¹¹, Nan Zhou²³, Shun Zhou¹¹,
 Tong Zhou¹¹, Xiang Zhou³⁴, Jingsen Zhu²⁹, Kangfu Zhu³⁵, Kejun Zhu¹¹, Zhihang Zhu¹¹,
 Bo Zhuang¹¹, Honglin Zhuang¹¹, Liang Zong¹⁴, Jiaheng Zou¹¹, Sebastian Zwickel⁵⁵

¹ Yerevan Physics Institute, Yerevan, Armenia² Université Libre de Bruxelles, Brussels, Belgium³ Universidade Estadual de Londrina, Londrina, Brazil⁴ Pontificia Universidade Católica do Rio de Janeiro, Rio de Janeiro, Brazil⁵ Millennium Institute for SubAtomic Physics at the High-energy Frontier (SAPHIR), ANID, Chile⁶ Pontificia Universidad Católica de Chile, Santiago, Chile⁷ Universidad Técnica Federico Santa María, Valparaíso, Chile⁸ Beijing Institute of Spacecraft Environment Engineering, Beijing, China⁹ Beijing Normal University, Beijing, China¹⁰ China Institute of Atomic Energy, Beijing, China¹¹ Institute of High Energy Physics, Beijing, China¹² North China Electric Power University, Beijing, China¹³ School of Physics, Peking University, Beijing, China¹⁴ Tsinghua University, Beijing, China¹⁵ University of Chinese Academy of Sciences, Beijing, China¹⁶ Jilin University, Changchun, China¹⁷ College of Electronic Science and Engineering, National University of Defense Technology, Changsha, China¹⁸ Chongqing University, Chongqing, China¹⁹ Dongguan University of Technology, Dongguan, China²⁰ Jinan University, Guangzhou, China²¹ Sun Yat-Sen University, Guangzhou, China²² Harbin Institute of Technology, Harbin, China²³ University of Science and Technology of China, Hefei, China²⁴ The Radiochemistry and Nuclear Chemistry Group in University of South China, Hengyang, China²⁵ Wuyi University, Jiangmen, China

- ²⁶ Shandong University, Jinan, China and Key Laboratory of Particle Physics and Particle Irradiation of Ministry of Education, Shandong University, Qingdao, China
- ²⁷ Nanjing University, Nanjing, China
- ²⁸ Guangxi University, Nanning, China
- ²⁹ East China University of Science and Technology, Shanghai, China
- ³⁰ School of Physics and Astronomy, Shanghai Jiao Tong University, Shanghai, China
- ³¹ Tsung-Dao Lee Institute, Shanghai Jiao Tong University, Shanghai, China
- ³² Institute of Hydrogeology and Environmental Geology, Chinese Academy of Geological Sciences, Shijiazhuang, China
- ³³ Nankai University, Tianjin, China
- ³⁴ Wuhan University, Wuhan, China
- ³⁵ Xi'an Jiaotong University, Xi'an, China
- ³⁶ Xiamen University, Xiamen, China
- ³⁷ School of Physics and Microelectronics, Zhengzhou University, Zhengzhou, China
- ³⁸ Institute of Physics, National Yang Ming Chiao Tung University, Hsinchu
- ³⁹ National United University, Miao-Li
- ⁴⁰ Department of Physics, National Taiwan University, Taipei
- ⁴¹ Charles University, Faculty of Mathematics and Physics, Prague, Czech Republic
- ⁴² University of Jyväskylä, Department of Physics, Jyväskylä, Finland
- ⁴³ IJCLab, Université Paris-Saclay, CNRS/IN2P3, 91405 Orsay, France
- ⁴⁴ University of Bordeaux, CNRS, LP2i, UMR 5797, F-33170 Gradignan, France
- ⁴⁵ IPHC, Université de Strasbourg, CNRS/IN2P3, F-67037 Strasbourg, France
- ⁴⁶ Aix Marseille Univ, CNRS/IN2P3, CPPM, Marseille, France
- ⁴⁷ SUBATECH, Université de Nantes, IMT Atlantique, CNRS-IN2P3, Nantes, France
- ⁴⁸ APC, Université de Paris, CNRS/IN2P3, F-75013 Paris, France
- ⁴⁹ LLR, École polytechnique, CNRS/IN2P3, F-91120 Palaiseau, France
- ⁵⁰ OMEGA, École polytechnique, CNRS/IN2P3, F-91120 Palaiseau, France
- ⁵¹ III. Physikalisches Institut B, RWTH Aachen University, Aachen, Germany
- ⁵² Institute of Experimental Physics, University of Hamburg, Hamburg, Germany
- ⁵³ Forschungszentrum Jülich GmbH, Nuclear Physics Institute IKP-2, Jülich, Germany
- ⁵⁴ Institute of Physics and EC PRISMA⁺, Johannes Gutenberg Universität Mainz, Mainz, Germany
- ⁵⁵ Technische Universität München, München, Germany
- ⁵⁶ Helmholtzzentrum für Schwerionenforschung, Planckstrasse 1, D-64291 Darmstadt, Germany
- ⁵⁷ Eberhard Karls Universität Tübingen, Physikalisches Institut, Tübingen, Germany
- ⁵⁸ INFN Catania and Dipartimento di Fisica e Astronomia dell'Università di Catania, Catania, Italy
- ⁵⁹ Department of Physics and Earth Science, University of Ferrara and INFN Sezione di Ferrara, Ferrara, Italy
- ⁶⁰ INFN Sezione di Milano and Dipartimento di Fisica dell'Università di Milano, Milano, Italy
- ⁶¹ INFN Milano Bicocca and University of Milano Bicocca, Milano, Italy
- ⁶² INFN Milano Bicocca and Politecnico di Milano, Milano, Italy
- ⁶³ INFN Sezione di Padova, Padova, Italy
- ⁶⁴ Dipartimento di Fisica e Astronomia dell'Università di Padova, Padova, Italy
- ⁶⁵ INFN Sezione di Perugia and Dipartimento di Chimica, Biologia e Biotecnologie dell'Università di Perugia, Perugia, Italy
- ⁶⁶ Laboratori Nazionali di Frascati dell'INFN, Roma, Italy
- ⁶⁷ University of Roma Tre and INFN Sezione Roma Tre, Roma, Italy
- ⁶⁸ Institute of Electronics and Computer Science, Riga, Latvia
- ⁶⁹ Pakistan Institute of Nuclear Science and Technology, Islamabad, Pakistan
- ⁷⁰ Joint Institute for Nuclear Research, Dubna, Russia
- ⁷¹ Institute for Nuclear Research of the Russian Academy of Sciences, Moscow, Russia
- ⁷² Lomonosov Moscow State University, Moscow, Russia
- ⁷³ Comenius University Bratislava, Faculty of Mathematics, Physics and Informatics, Bratislava, Slovakia
- ⁷⁴ Department of Physics, Faculty of Science, Chulalongkorn University, Bangkok, Thailand
- ⁷⁵ National Astronomical Research Institute of Thailand, Chiang Mai, Thailand
- ⁷⁶ Suranaree University of Technology, Nakhon Ratchasima, Thailand
- ⁷⁷ Department of Physics and Astronomy, University of California, Irvine, CA, USA

ARTICLE INFO

Keywords:

JUNO
Top Tracker
Plastic scintillator
Photomultiplier
PMT
Multianode
WLS fibre
Neutrino
Muons

ABSTRACT

The main task of the Top Tracker detector of the neutrino reactor experiment Jiangmen Underground Neutrino Observatory (JUNO) is to reconstruct and extrapolate atmospheric muon tracks down to the central detector. This muon tracker will help to evaluate the contribution of the cosmogenic background to the signal. The Top Tracker is located above JUNO's water Cherenkov Detector and Central Detector, covering about 60% of the surface above them. The JUNO Top Tracker is constituted by the decommissioned OPERA experiment Target Tracker modules. The technology used consists in walls of two planes of plastic scintillator strips, one per transverse direction. Wavelength shifting fibres collect the light signal emitted by the scintillator strips and guide it to both ends where it is read by multianode photomultiplier tubes. Compared to the OPERA Target Tracker, the JUNO Top Tracker uses new electronics able to cope with the high rate produced by the high rock radioactivity compared to the one in Gran Sasso underground laboratory. This paper will present the new electronics and mechanical structure developed for the Top Tracker of JUNO along with its expected performance based on the current detector simulation.

1. Introduction

The Jiangmen Underground Neutrino Observatory (JUNO) [1,2], under installation in the South of China at a distance of about 53 km from two nuclear plants, is a multipurpose experiment with the main goal of determining the neutrino mass ordering. For this purpose, the

experiment will use a 20 kt liquid scintillator detector with a ~3% energy resolution at 1 MeV. At the same time, JUNO will also measure with a sub-percent accuracy the oscillation parameters θ_{12} , Δm_{21}^2 and Δm_{31}^2 . JUNO will be able to detect neutrinos from other natural sources as, e.g., geo-neutrinos and supernova explosions [2,3].




Fig. 1. Schematic view of the JUNO detector with the (a) central detector, (b) water Cherenkov detector and (c) Top Tracker. The (d) calibration house and the chimney connecting it to the central detector are also shown. Configurations considered for the induced noise, from either fast neutrons or cosmogenic isotopes, in the central detector, the water Cherenkov detector and the (e) surrounding rock are also illustrated.

The JUNO detector is composed of a central detector containing an acrylic sphere in which the liquid scintillator is located, surrounded by 17 612 large (20") and 25 600 small (3") photomultipliers (PMTs). The central detector is immersed in a water Cherenkov detector instrumented with 2400 large PMTs. A muon tracker, called Top Tracker (TT), is placed on top of those detectors. The water Cherenkov detector and the Top Tracker combined form the Veto system of JUNO that was conceived to track and veto almost all muons crossing the detector. A detailed description of the experimental equipment can be found in [2] and the above listed elements are also shown in Fig. 1. In this figure is also identified the position of the calibration house, required for the central detector calibration, and the chimney connecting them. In order to fit the chimney and calibration house, the central layers of the Top Tracker were moved to a higher position, as is also shown in Fig. 1. The detector will be placed in an underground laboratory with an overburden corresponding to about 700 m depth (1800 m.w.e). At this depth it is still expected to have about 15 atmospheric muons per hour and per m^2 with a mean energy of 207 GeV. These muon tracks can induce a background very similar to the inverse beta decay (IBD) interactions produced by reactor neutrinos.

The most dangerous background induced by these atmospheric muons, which are produced by the interaction of cosmic-rays with the Earth's atmosphere, comes from the generation in the detector or in the surrounding rock of ${}^9\text{Li}$ and ${}^8\text{He}$ unstable elements, and fast neutrons. Fig. 1 presents schematically these background contributions. The main contribution comes from ${}^9\text{Li}$ and ${}^8\text{He}$ produced directly in the central detector and in the water Cherenkov detector while the last one concerns the neutron production in the surrounding rock.

The simulations show that IBD interactions and cosmogenic background mimicking these interactions have comparable rates. For this reason the rejection of this background contribution and the study of its impact on the systematic errors is very important in JUNO. This is the main role of the Top Tracker. All OPERA experiment [4] Target Tracker [5] plastic scintillating modules have been recovered and sent to China where they will be used to form the JUNO Top Tracker.

In the OPERA experiment, the Target Tracker modules were shielded by the emulsion/lead bricks, while in the JUNO cavern the TT will be directly exposed to rock radioactivity. On top of that, the radioactivity in the JUNO cavern is measured to be about two orders of magnitude higher than the one observed in the underground Laboratori Nazionali del Gran Sasso (LNGS) [6] where the OPERA experiment was located. For these reasons, the electronics of the OPERA Target Tracker had to be replaced by a significantly faster version.

The 496 OPERA Target Tracker modules are enough to cover, with at least 3 layers of TT walls, only about 60% of the top surface of the JUNO central detector and thus can track only 30% of the total muon flux. For this reason, the TT will participate in the JUNO veto for the muons traversing it. It is worth noticing here that about half of the atmospheric muons pass by the sides of the water Cherenkov detector without crossing the top surface where the TT is located. The water Cherenkov detector surrounding all the JUNO central detector will be used as veto to reject events coming from all directions, including those missed by the TT.

The ${}^9\text{Li}$ and ${}^8\text{He}$ candidate events identified using a correlation between an IBD-like signal and the TT muon reconstruction will make it possible to create a data driven estimate of the energy spectrum and rate of this background to be used in analysis. The specific characteristics of this sample are still being evaluated. The TT can also be used to select and reconstruct a sample of muon tracks to be used to tune and validate the reconstruction algorithms based on central and water Cherenkov detectors' information. This is possible because of the centimetre level accuracy on the position of the muon tracks crossing the TT.

This paper describes in Section 2 the detection principle and TT geometry. This is followed by a description of the expected rates in the detector along different levels of trigger in Section 3. The Top Tracker electronics designed to perform the data readout and trigger under the expected background conditions is extensively presented in Section 4. In Section 5 the current Top Tracker reconstruction and its performance is presented. Details on the operation modes of the Top Tracker are detailed in Section 6, which is followed by a description of a prototype of the Top Tracker in Section 7 where these modes were extensively tested. Elements concerning the supporting structure for the Top Tracker modules and installation are provided in Section 8. Long term stability monitoring of the Top Tracker modules is presented in Section 9. Finally, a summary and concluding remarks are presented in Section 10.

2. Description of the top tracker

As described in [5] presenting the OPERA Target Tracker, the TT uses 6.86 m long, 10.6 mm thick, 26.4 mm wide scintillator strips¹ read on both sides using Kuraray² Y11(175)-MJ Wave Length Shifting (WLS) fibres and Hamamatsu³ 8×8 channels multianode photomultipliers (MA-PMT) H7546. The technology used in the TT is very reliable due to the robustness of its components. Delicate elements, like electronics and MA-PMTs are located outside the sensitive area where they are easily accessible.

A basic unit (module) of the TT consists of 64 strips read out by WLS fibres routed at both ends to two 8×8 -channel photodetectors placed inside the end-caps (Fig. 2). Four modules are assembled together to construct a tracker plane covering $6.7 \times 6.7 m^2$ sensitive surface. Two planes of 4 modules each are placed perpendicularly to form

¹ Composed of polystyrene with 2% of p-Terphenyl and 0.02% of POPOP. Produced by AMCRYS-H, 60, Lenin ave, Kharkov, 310001, Ukraine.

² Kuraray Co., Methacrylic Resin Division, 8F, Maruzen Building, 3-10, 2-Chrome, Nihonbashi, Chuo-ku, Tokyo, 103-0027, Japan.

³ Hamamatsu Photonics K.K., Electron Tube Center, 314-5, Shimokanzo, Toyooka-village, Iwata-gun, Shizuoka-ken, 438-0193, Japan.




Fig. 2. Schematic view of an end-cap of a scintillator strip module.




Fig. 3. Schematic view of a plastic scintillator strip wall. In addition to the 4 modules clearly visible in the drawing, there are 4 additional modules placed below them in a perpendicular orientation, going for example between the PMTs numbered 0 and 11.

a tracker wall providing 2D track information (Fig. 3). The inherited OPERA Target Tracker contains 62 walls, with a total number of 31 744 scintillating strips for 63 488 electronic channels.

The JUNO Top Tracker walls, contrarily to their utilisation in OPERA, have to be placed horizontally as shown by Fig. 4. A bridge is placed on top of the JUNO Central Detector serving as support for the TT structure. The three layers of the TT are placed on this bridge, with overlapping walls to limit the dead zones. The distance

Table 1

Rock radioactivity in LNGS and in the JUNO site.

| Isotope | Activity (Bq/kg) | |
|-------------------|------------------|---------------|
| | LNGS | JUNO site |
| ^{40}K | 26 ± 2 | 1340 ± 50 |
| ^{238}U | 1.8 ± 1 | 110 ± 10 |
| ^{232}Th | 1.5 ± 1 | 105 ± 10 |

of 1.5 m between each layer is a compromise between the muon track reconstruction accuracy and the total size of the needed supporting mechanical structure. The position of the bottom layer is defined by the minimum distance between the surface of the water of the Cherenkov veto detector and the TT. Three walls over a distance of 20 m are placed in one horizontal direction and 7 walls over a distance of 47 m in the other direction, leaving a hole in the middle for the JUNO chimney (20 walls per layer). In this way, 60 walls are used, while the modules of the remaining two walls are reshuffled and placed on top of the chimney, part of the JUNO detector not surrounded by the Water Cherenkov veto detector. To have three full layers just on top of the chimney, one full TT wall is placed in the lowest part and two partial TT walls are placed in the middle and upper parts, each of those separated by 23 cm. The partial walls above the chimney contain only the 2 central modules in each layer, for a total of 4 modules per wall, and thus fully covering only the centre of the chimney.

According to their position, the modules of each wall are placed facing the top or bottom directions in a way to have their end caps, and, in particular, their electronics, accessible from outside.

3. Trigger rate

While, as discussed previously, the goal of the TT is to detect atmospheric muons, these are not the most abundant particles reaching the TT. Indeed, the natural radioactivity of the elements composing the TT (0.29 ± 0.05 Hz/strip, or ~ 9 kHz in the entire detector) [5] is larger than the expected ~ 4 Hz of muons crossing the entire Top Tracker. This natural radioactivity rate is, however, also about three orders of magnitude smaller than the events induced in the TT due to the radioactivity of the rock of the cavern where the detector will be installed. The trigger rate of the TT will be directly tied to this rock radioactivity, which will be described in this section. To cope with it, specific electronics described in Section 4 will be needed. Due to the sizeable difference in rate between the detector natural radioactivity and that from the cavern, along with an expected similarity between the two types of signature, the detector's natural radioactivity will be neglected for the rest of this paper.

In addition to the radioactivity, the PMT dark count rate will also affect the rate of triggered events. The mean dark count rate of the channels of the MA-PMT used in the TT is 2.45 Hz (corresponding to ~ 78 kHz in the entire detector) [5]. This rate is also significantly lower than the one expected due to rock radioactivity and, therefore, it was also neglected for the rest of this paper.

During the prospecting of the JUNO site, rock samples were collected to determine the content of radioisotopes. It was determined⁴ that the rock radioactivity at the JUNO site is about two orders of magnitude higher than that in the LNGS where the detector had been previously used, as shown in Table 1. In addition to the higher radioactivity from the cavern, in JUNO the TT will also be directly exposed to the rock, while the OPERA TT modules were shielded by lead/emulsion bricks.

The MA-PMT trigger rate induced by the rock radioactivity is estimated using the JUNO official simulation software [7], where the

⁴ Measurements realised by Ph. Hubert (CENBG Bordeaux) using a Ge detector.




Fig. 4. Schematic view of the JUNO Top Tracker on top of the Central Detector.




Fig. 5. Trigger rate per MA-PMT for the three TT layers. The vertical axis represents how many MA-PMT have each expected average rate shown in the horizontal axis. A MA-PMT trigger threshold of $1/3$ p.e. is assumed.

TT modules are placed as previously described. In this study decays were individually simulated from the surrounding rock (within a 50 cm thickness from the rock surface) and the decay particles are then propagated using GEANT4 [8]. All energy deposited in the plastic scintillator is then converted into the expected signal at the MA-PMT based on the detector description in [5]. The MA-PMT is triggered when the collected signal is higher than the threshold of $1/3$ photoelectron (p.e.). Based on these simulations, the average trigger rate per MA-PMT varies between 20 kHz and 60 kHz, as shown in Fig. 5. The trigger rate per channel has an average value of about 670 Hz, with some channels exceeding 1 kHz, which has to be compared with about a 20 Hz rate for the TT operation in the LNGS. As is also shown in Fig. 5, the rate observed in each MA-PMT will depend on its position given that the modules closer to the rock will be more exposed to radiation than the more central ones, and the upper layers provide some shielding for the lower ones.

The high trigger rate at the MA-PMT level expected due to the rock radioactivity imposes strong constraints on the new TT acquisition system. Given that the particles emitted by radioactivity produced by the decay of ^{40}K , ^{232}Th , ^{238}U and their decay products will not typically traverse more than one layer of scintillator, it is possible to reduce the amount of data generated during the JUNO experiment by requiring a time coincidence between triggers related to MA-PMTs along perpendicular modules in the same wall. Additionally, to further reduce the produced data, it is also required that these time coincidences are found in the three aligned TT walls of different layers. With these coincidence criteria the full detector event rate can be reduced from about 8 MHz, which corresponds to the total rate over the full detector from the per PMT rates shown in Fig. 5, to about 2 kHz, as shown in Fig. 6, which can be handled by the acquisition system. The two levels of time coincidence described correspond to the L1 and L2 triggers. The L1 trigger relates to the time coincidences within a single wall while the L2 trigger does the same at the level of the entire TT. The implementation of these coincidence criteria in the electronics is discussed in the following sections.

It is worth noting that in Fig. 6 the L1 and L2 algorithms that are the most likely to be used in the Top Tracker were chosen, however the final decision has not yet been made. The L1 trigger, which is discussed in Section 4.3, provides about an order of magnitude reduction of the rate by requiring 100 ns coincident MA-PMT triggers in the two perpendicular planes of any given wall, and with at least one of the planes having the MA-PMTs on both sides of the module triggered. The L2 trigger, which is discussed in Section 4.4, provides another reduction factor of about 400 to the event rate by requiring at least 3 aligned walls on different layers of the detector to have produced a L1 trigger within a 300 ns sliding time window. In case an alternative algorithm is used in the L1 and L2 electronics cards, due to difficulty in implementing them, and the data reduction is not as efficient as discussed here, it is foreseen that the data acquisition system will be able to apply online additional criteria to reduce it to the rate described above.

It is also useful to note, as shown in Fig. 6 that the radioactivity rate can be further reduced offline by performing a full 3D reconstruction of the muon track, rather than just requiring walls to be aligned. This




Fig. 6. Total event rate of the full detector at various trigger levels, from the MA-PMT trigger rate up to the rate after events are reconstructed by the Top Tracker offline software, due to radioactivity in the surrounding rocks (green) and due to atmospheric muons (blue). The L1 selection criteria requires an X and Y coincidence involving 3 out of 4 MA-PMTs at the end of the corresponding modules in the perpendicular X and Y directions. The L2 selection criteria requires the L1 coincidences to be on 3 walls of the Top Tracker located on different layers.

reconstruction is presented in Section 5 and corresponds to the “offline reco” level on the aforementioned figure, for which only a 68% upper limit is shown for radioactivity. Any potential remaining radioactivity induced events after the reconstruction can still be further reduced by requiring a correlated signal between the Top Tracker and the central detector or the water Cherenkov veto detector, after that the impact of the radioactivity becomes negligible compared to the atmospheric muon rate.

4. Electronics

The global electronics conceptual readout chain is shown in Fig. 7. As it was previously mentioned, the TT is a set of 63 identical walls with similar mechanical constraints per wall which leads to the development of individual electronics per wall. This electronics chain is implemented at the TT wall level as shown in Fig. 8. Due to the limited thickness of the module end-caps (3 cm) the PMT readout had to be split in two boards: the front end board (FEB) directly connected to the MA-PMTs and the readout board (ROB) connected to the FEB and the wall concentrator board (CB). Both FEB and ROB are placed in the module end-caps next to the MA-PMT. The CB, located in the centre of each wall, configures and controls the 16 FEBs and ROBs in a wall, and performs the data readout for the full wall which will subsequently be transferred to the data acquisition system by the CB. The MA-PMTs are read out by the 64-channel MAROC3 [9] ASIC placed on the FEB. The ROB is placed after the FEB to receive the MA-PMT data and to send them to the CB that gathers the data from the 16 sensors of a TT wall. A split-power board placed at the level of the CB powers the MA-PMTs, the FEB, the ROB and the CB.

The charge and hit information are given by the MAROC3 chip together with an “OR” trigger signal used by the CB for time-stamping. The CB identifies x-y correlations at the level of the wall (L1 trigger) with a sliding time window of 100 ns. In the absence of correlations in a wall the CB will send back a RESET signal to the ROB in order to reduce the dead time of the FEB MAROC3 (between 7–14 μ s). In case a coincidence is detected the CB sends this information to a Global Trigger Board (GTB), which further searches for correlations at the level of all three TT layers (L2 trigger) with a sliding window of 300 ns. The GTB sends back a validation or a rejection signal to the CBs that, correspondingly, either sends the related data to the JUNO data acquisition system or issues a reset signal to the relevant ROB in about 1 μ s. Both trigger levels are essential to significantly reduce the TT data rate, as discussed previously, and to reduce the detector dead time of the FEB MAROC3.




Fig. 7. TT electronics chain. All electronics cards developed for the TT are shown in blue and are the front end board (FEB), readout board (ROB), concentrator board (CB) and global trigger board (GTB). The FEB and ROB are installed in each module end-cap. The FEB is connected to the MA-PMT that observes the light collected by WLS fibres in the 64 scintillator strips of each module and to the ROB. The ROB is connected to the FEB and to the CB of that specific wall. The CB is located at the centre of the wall and is connected to all 16 ROBs of the wall. The CB is also connected to the white rabbit (WR) switch, to the data acquisition system (DAQ) and to the GTB. The GTB, in addition to being connected to the 63 CBs of the TT, will also be connected to the WR switch.




Fig. 8. Schematic view of the TT wall electronics. The FEB and ROB at each end-cap are connected to the CB which performs the L1 trigger. The CB connections are shown as boxes in the top right corner of the CB box. In particular the CB is connected to the L2, to the white rabbit (WR) switch via UDP and to the data acquisition system via TCP. Next to the CB in the wall is also located the split-power board that distributes the power for all ROBs and the CB in each wall.

4.1. Front End Board (FEB)

The FEB is an 8-layer PCB, 3 cm \times 13 cm in size, carrying the MAROC3 and a Spartan6 XC6SLX4 FPGA used to serialise the hit signals from 64 parallel outputs to 8 serial links. This operation allows the ROB to obtain the information of which channels have triggered, which is then used for online zero suppression if that is the set acquisition mode. The board is directly plugged to the MA-PMT, as shown in Fig. 9. The lines from the MA-PMT to the MAROC3 inputs are separated by at least twice the PCB track width (0.15 mm) to minimise coupling between tracks, and protected from external noise sources by four ground planes in the PCB.

The readout electronics of the Top Tracker is based on a 64-channel ASIC, the MAROC3 designed by the Omega laboratory [9]. The chip is provided in a 12 mm \times 12 mm TFBGA package with 353 pins. The main characteristics of this chip are:




Fig. 9. The FEB connected to a MA-PMT.

- Configurable gain compensation (called gain equalisation factor), varying from 0 to 3.9, to equalise the anode-to-anode gain differences of the MA-PMT channels [5]. The MAROC3 is equipped with an adjustable gain system incorporated in the preamplifier stage delivering an identical amplification factor to both, the fast shaper (trigger arm) and slow shaper (charge measurement arm), of each channel.
- Delivery of a global low noise auto-trigger with 100% trigger efficiency for minimum ionising particles, using a threshold as low as $1/3$ p.e., corresponding to 53 fC for a total gain of 10^6 .
- Delivery of a charge proportional to the signal detected by each MA-PMT channel in a dynamic range up to 5 pC (~ 30 p.e.⁵).

As mentioned in the list above, each of the 64 channels comprises a low noise variable gain preamplifier that feeds both a trigger and a charge measurement arm (Fig. 10). The auto-trigger includes a fast shaper followed by a comparator. This signal is compared with an adjustable threshold to deliver a trigger signal for each channel. The trigger decision is provided by the logical “OR” of all 64 channels. A malfunctioning channel can be disabled from the trigger decision through a mask register.

The charge measurement arm consists of a slow shaper followed by a Track&Hold buffer. The slow shaper has a gain of 121 mV/pC (equivalent to 19 mV/p.e.) and a long peaking time to minimise the sensitivity to the signal arrival time. Upon a trigger decision, charges are stored in capacitors and the 64 channels are readout sequentially through a single analogue output, called OutQ, at a 10 MHz frequency. The charge output can also be directly read with an 8, 10 or 12 bits internal Wilkinson ADC. For the TT the 8 bit readout was selected to be used with the Wilkinson ADC to reduce the readout deadtime. The output clocking frequency is 80 MHz. In the 8 bit case all 64 channels are read in less than 14 μ s, and while the channels are being read the MAROC3 will not be available to measure another charge resulting in detector deadtime. This time can be reduced if a reset signal is received from the CB via the ROB, at which point the readout and transfer process is stopped and the MAROC3 is ready to start again charge measurement. It is worth noting that this deadtime does not affect the trigger arm of the MAROC3 discussed in the previous paragraph.

The MAROC3 chip consumes approximately 220 mW (3.5 mW/channel), with a small variation depending upon the gain correction settings and ranges.

The variable gain system is activated by eight switches which allow setting the gain equalisation factor varying from 0 to 3.9. By turning off all current switches of a channel, the gain equalisation factor is zeroed and the channel is disabled. For a gain equalisation factor of 1, the switches are set to 64.

The trigger efficiency can be measured as a function of the injected charge for each individual channel. 100% trigger efficiency is obtained

for input charge as low as 0.1 p.e., independently of the gain equalisation factor. With the trigger threshold set externally and common to all channels, the output spread among the 64 channels can be carefully controlled. It is found to be around 0.01 p.e., an order of magnitude smaller than the useful threshold level.

Upon a trigger decision, all the capacitors storing the charge are read sequentially through a shift register made by D-flip-flop. The pedestal level is less than $1/3$ p.e. considered as the final TT trigger level.

The reproducibility in the charge measurement has been determined for all channels of the same chip and found to be better than 2% over the full range 1–10 pC for a gain equalisation factor of 1, corresponding to 1–35 p.e. (Fig. 11).

Cross-talk related to the ASIC has been carefully considered. Two main sources of cross-talk have been identified. The first one comes from a coupling between the trigger and the charge measurement arms and has been determined to be lower than 0.1%. The second one affects the nearest neighbours of a hit channel, where a cross-talk of the order of 1% has been measured. This effect is mainly due to the routing of traces in the board and is negligible for far channels.

The FEB contains buffer amplifiers for the differential charge output signals of the MAROC3 and logic level translators for the digital signals. The FEB is connected to the ROB board by means of one 100 pins coaxial connector. This two-board arrangement conforms to the mechanical constraints of the TT module’s end-cap geometry.

4.2. ReadOut Board (ROB)

The ROB manages the FEB power, acquisition and slow control. It is an 8-layer PCB developed in collaboration with CAEN (mod. A1703) based on an ALTERA Cyclone V GX FPGA. A picture of the PCB top and bottom views is shown in Fig. 12.

In standard acquisition mode, the FPGA state machine waits for the detection of a FEB “OR” signal. Following the detection of this signal, a Hold signal is issued, for charge measurement in the MAROC3 chip, with a tunable delay and a time jitter lower than 1 ns. In this way, a charge measurement precision better than 1% is achieved. The triggered channel pattern is also acquired and used for zero-suppression, as discussed previously. The ROB features a 12 bit Flash ADC converter AD9629BCPZ-65 [10] by Analog Devices for the read out of the OutQ signal from MAROC3 chip (analogue output proportional to the charge to be measured). This Flash ADC converter samples the OutQ signal at 20 MHz, which corresponds to two samples for each different OutQ value. In such a way, the readout of one event is performed in about 7 μ s. The readout of the charge converted by the internal Wilkinson ADC discussed previously is also possible, however that will require up to the previously mentioned 14 μ s. Due to this difference in the time required for the charge readout, the Flash ADC is expected to be used for charge measurement during data taking.

The ROB also hosts a High Voltage module to supply the MA-PMT, which is operated in a range from -1000 to -800 V, with a current lower than 500 μ A, and a test pulse unit for calibration purposes.

The HV module is the A7511N DC–DC converter by CAEN [11] with a ripple lower than 10 mV. Tests carried out on a detector prototype have shown that the DC–DC converter noise contribution to the MA-PMT output signal is lower than 1 mV.

The test pulse generators, illuminating the fibres in the end-caps of the TT modules, are the same used in OPERA [5]. The generator produces short pulses, 20 ns wide with adjustable amplitude, powering blue LEDs which excite the WLS fibres exiting the scintillator strips. In the calibration procedure the pulse amplitude is set in such a way to simulate the single photoelectron signal on the MA-PMT while the HOLD signal is issued with an adjustable delay with respect to the LED driver pulses, emulating the acquisition of standard signals. Pedestals can be measured with random emission of HOLD signals inside the

⁵ A gain of 10^6 is assumed in all charge to p.e. conversions in this paper.




Fig. 10. Architecture of MAROC3 ASIC [9].




Fig. 11. Linearity of the charge measurement as function of the input charge for a gain equalisation factor set to 0.5, 1, 2 and 3.

ROB. These various data acquisition modes will be further described in Section 6.

At nominal working conditions on a detector prototype, an operating current of 350 mA has been measured at the nominal operating voltage of 12 V supplied to the ROB, powering also the FEB and the MA-PMT.

Finally, the ROB includes the possibility to remotely re-flash the FEB Spartan6 FPGA, to facilitate remote system operation.

4.3. Concentrator Board (CB)

At the centre of each TT wall the CB is located, as shown in Fig. 8. Its main role is to aggregate the 16 incoming ROB data streams, perform data selection (Level 1 trigger), and forward the reduced data stream to the data acquisition system. All 63 CBs present in the detector operate in parallel and fully independently, but with a synchronised timing distribution system based on the White-Rabbit (WR) standard [12].

4.3.1. Interfacing to 16 ROB

Placing the CB at each wall makes it possible to reduce the number of connections to the data acquisition system from 992 (counting one for each ROB) to just 63 for the whole detector. Such wall-centric organisation is shown in Fig. 8.

The board provides four remote connections: one copper cable for the system power supply at +12 V (PWR) and three optical fibres for long-distance data links. One of the optical links implemented in the CB is used for the connection to the Detector Control System using the User Datagram Protocol (UDP) and for the WR time synchronisation system. Another optical link implements a dedicated ethernet connection (TCP) to the data acquisition system and the final link connects the CB to the GTB (see Section 4.4) for the Level 2 trigger selection.

All local connections between ROB and the CB are implemented using Cat5e cables. Associated with the CB board is a power-distribution card referred to as Split-Power Board that is briefly described in Section 4.3.4.

4.3.2. L1 trigger selection

As explained in Section 3, the data stream produced by a ROB card is dominated by radioactivity induced background events. At the expected




Fig. 12. Top and bottom sides of the ReadOut Board (ROB).

~50 kHz PMT trigger rate, the resulting data rate from a thousand ROBAs would be unmanageable at the data acquisition level.

As discussed previously, a reduction of the data rate by an order of magnitude can be achieved using the CB to select only events in x-y coincidence (L1 trigger level). A coincidence is defined here as a set of 2, 3 or 4 FEB/ROB triggers arriving in the same time window of 100 ns and associated with two perpendicular modules in the same wall. The L1 selection algorithm is implemented as a sliding window on the signals from all 16 ROBAs. To protect from classification ambiguities, any coincidence detected within this window will result in accepting all ROB data inside the window, even if some signals are not correlated to the event.

The triggers arriving at the CB are time-stamped with a resolution of 1 ns using the absolute time reference provided by the WR node implemented in the CB. For FEB/ROB triggers found to be in coincidence, the CB waits for the associated ROB data packets containing the signal position and its charge value. After merging the timing and charge information, the modified packets are transferred to the data acquisition system, provided the L2 trigger did not reject this L1 trigger. FEB/ROB triggers that do not produce a L1 trigger, or that are rejected by the L2 trigger, are rejected and a reset signal is sent to the corresponding ROB to abort charge digitisation as well as the associated data transfer to the CB.

The overall system synchronisation provided by the WR protocol is necessary for a correct data reconstruction in the data acquisition system and for data analysis. It is also used for communicating the coincidences to the GTB (a partial timestamp of the L1 trigger is sent from the CB to the GTB) that will confirm or reject them in accordance with the L2 trigger selection applied at the detector level.

4.3.3. CB implementation

The CB has been implemented as an FPGA-centred readout system with a motherboard that accommodates a daughtercard hosting the processing unit (System-on-Module, SoM). A block diagram of the CB is shown in Fig. 13 and the pre-production prototype is shown in Fig. 14.

The CB motherboard is a 34 cm × 20 cm, 8-layer card that provides all the interface-specific electronics components for supporting the LVDS⁶-based connections to the 16 ROBAs and the optical connections




Fig. 13. Concentrator Board (CB) block diagram.




Fig. 14. Concentrator Board (CB) motherboard and the daughtercard.

as described earlier. It provides a set of all necessary power supply voltages and their sequenced activation to satisfy the requirements of this FPGA-based readout system. The board also hosts an integrated WR-node that is implemented in a Spartan6 FPGA [13]. This FPGA also interfaces the slow-control UDP traffic going to and from the CB with the main FPGA in the daughtercard.

The daughtercard is an embedded high-performance, FPGA-based system on module, which controls and processes all incoming and

⁶ Low Voltage Differential Signalling.




Fig. 15. Split-power board.

outgoing data streams. It is a 10 cm × 7 cm, 16-layer PCB that hosts the Xilinx System-on-Chip (SoC) of the latest generation, Zynq Ultra-scale+CG [14], accompanied by a 2 GB of on-board DDR4 memory. The programmable-logic part of this FPGA controls the communication with the 16 ROBs and the link to the GTB, both based on custom protocols. Two of the four available dedicated hardware CPUs in the SoC manage in parallel the TCP traffic going to the data acquisition system and the slow-control.

4.3.4. Split-power board

The Split-power board, shown in Fig. 15, is a dedicated PCB for distributing power supply to all TT wall components. This board is approximately 15 cm × 15 cm in size and implemented in 2 copper layers. The power delivered by a remote power supply is fanned out to 16 ROBs and one CB. The board features re-settable fuses, shunt resistors and 16-bit precision I2C⁷ current/voltage monitors in all the output power branches. An I2C bus runs between this PCB and the CB to allow the CB to monitor the power consumption of all these readout components.

4.4. Global Trigger Board (GTB)

The main role of the GTB is to perform a second level (L2) trigger working on the x–y coincidences selected by the L1 trigger of each TT wall. The idea is to accept interesting events depending on the spatial position of the walls involved by the passage of the crossing muons. The requirement of the algorithm implemented in the GTB is to find an alignment of the x–y coincidences over the three TT layers in a given time window. The second requirement is that this algorithm has to be very fast in order to give an online feedback. As discussed previously, this level of trigger will be able to further remove the background noise given by radioactivity, correlated electronic noise or MA-PMT dark noise.

The GTB will occupy the top of the trigger chain acting as the final sink of all the trigger requests and the single source of all the trigger validations. Such topmost position will give the possibility for the GTB to achieve also secondary tasks typical of a trigger root node, such as an injection of a technical trigger (periodically or randomly generated), a throttling of the trigger in case of anomalous situation in which part of the detector is affected by more noise, or a prescale of the trigger for commissioning studies. GTB will have to take validation/rejection decisions based only on the spatial position of the incoming requests and on the trigger time information associated to them. Indeed, each




Fig. 16. Schematic view of the TT global trigger strategy.

trigger request will be associated with a timestamp that specifies the time the corresponding hits have been generated in the FEB.

Currently, the algorithm for finding aligned x–y correlations over the three TT layers has not been decided yet. Several solutions have been studied in simulation. The most promising is the coincidence of three aligned walls each on a different layer, however the possibility of just requiring three walls without any alignment was also considered. Fig. 6 shows the reduction in terms of the total event rate applied by the L2 trigger for the most promising case. When only three walls, without any alignment, are required the GTB performs about 20 times worse. This both underlines the importance of alignment, while still showing that even a significantly less resource demanding algorithm on the GTB would already provide more than an order of magnitude reduction to the radioactivity rate passing the L1 trigger. Further solutions exploiting neural networks algorithms implemented directly on the FPGA are also under study.

The hardware implementation of the GTB relies on a single device connected to all the CBs (see Section 4.3) by means of bidirectional optical links. Indeed, each CB should send the trigger requests and receive back a validation/rejection that will be used for resetting the FEB and the data acquisition buffers. Fig. 16 shows a sketch of the connection scheme where the GTB is placed in the middle of the detector. The GTB should also accommodate a connection to the WR network that would guarantee the synchronisation with the CBs and will be used for configuring the GTB. An additional optical link is also needed for the GTB to be able to receive external triggers from the other JUNO systems. A total of 65 optical links is therefore needed in the GTB as shown in Fig. 17. While these two latter links are Gb connections, the 63 links from the CB are not expected to have a bandwidth larger than 400 Mbps since only a partial timestamp will be sent along with the trigger requests. This is possible because the GTB belongs to the common clock distribution system and hence it receives its full timestamp from the WR network.

The GTB is implemented as a motherboard hosting a SoM. Fig. 17 shows a block diagram of the GTB motherboard and Fig. 18 shows a photo of the board. It will accommodate the optical modules and provide the power needed by the SoM that will host an FPGA able to receive the two network links on fast transceivers and the 63 CB links on normal pins in order to minimise the latency of the link. For homogeneity with the rest of the system, a Xilinx FPGA has been chosen. In particular, the commercially available off-the-shelf module Origami B20 [15] has been chosen as SoM for the GTB. It hosts a Kintex Ultrascale XCKU060 FPGA [16] that has more than 150 single ended I/Os routed to a Z-Ray connector.

5. TT event reconstruction

The main goal of the TT reconstruction is to determine the direction of atmospheric muons crossing the detector. In order to achieve this goal, the track reconstruction is divided in three successive steps. In the first step, for each MA-PMT, neighbouring triggered channels are clustered to reduce the impact of the cross-talk in the reconstruction.

⁷ Inter-Integrated Circuit.




Fig. 17. Block diagram of the GTB motherboard. The GTB motherboard hosts a SoM, a White Rabbit (WR) endpoint for clock distribution and synchronisation and 65 bidirectional optical transceivers (SFP+).




Fig. 18. Global Trigger Board (GTB) motherboard and SoM.

In the second step, the 3D crossing points of the muon are reconstructed based on information from both perpendicular detection planes of each wall participating in the x-y coincidence. In the final step, a 3D line is fitted to these 3D points via a χ^2 minimisation. This last step is done independently for every ensemble of more than three 3D points at different vertical positions, and fits with a $\chi^2 \geq 3$ are rejected. In case only three 3D points remain in the fit, the χ^2 criteria is relaxed to accept all fits with $\chi^2 < 10$.

This track reconstruction is able to reduce the remaining radioactive noise rate in the detector by at least 2 orders of magnitude, as shown in Fig. 6, achieving a rate comparable to the expected atmospheric muon rate in the TT. While this algorithm is very efficient to reject background events due to radioactivity, it is also able to reconstruct almost all muons passing through the detector, with $\sim 97\%$ of the muons having passed the L2 trigger condition being reconstructed in a sample simulated merging individual muons⁸ and the expected radioactive noise in a $\pm 15 \mu\text{s}$ time window around them.

Thanks to the high granularity of the TT and the lever arm provided by the distance between its layers, the median angular resolution of the detector is of about 0.2° , as shown in Fig. 19. Projecting both the simulated and reconstructed tracks to the bottom of the water Cherenkov

⁸ The muons were simulated using the expected energy and angular distribution in the JUNO cavern.




Fig. 19. Expected angular resolution of the TT reconstruction. The last point includes also all events having an angular resolution larger than 2° .




Fig. 20. Distance between the points defined by the crossing of the plane corresponding to the bottom of the water Cherenkov detector (WCD) and the lines from the simulated and reconstructed muons. Note that the bottom of the WCD is 45 m below the bottom of the TT, and corresponds to the largest projected error of the TT tracking within JUNO. The last point includes also all events with a distance larger than 2 m.

detector, which corresponds to the largest error of the reconstruction while still within the JUNO detector, the median resolution would correspond to about 20 cm, as shown in Fig. 20.

The long tail in these resolutions, containing about 1% of all reconstructed events, is due to poorly filtered cross-talk and noise leading to reconstructed tracks not corresponding to the simulated muons. Some more strict quality criteria for the TT reconstruction are under evaluation which can significantly reduce this tail, however at the moment they also reduce by a third the fraction of reconstructed muons, which is not desirable. Alternative reconstruction algorithms are also under evaluation. Going beyond looking only at TT data, a joint reconstruction using the TT, the water Cherenkov detector and the central detector is expected to be able to easily identify all cases where the TT alone identified a fake muon track and thus reduce both the long tail in the TT resolution and remove any remaining noise-induced events in the TT, without negatively impacting on the reconstruction efficiency. This joint reconstruction is still under study.

6. TT operation modes

There are four different modes foreseen for the TT operation. These modes were thought to make it possible to monitor the TT stability,

Table 2
TT operation mode characteristics.

| | TRT | PED | LED | Normal |
|------------------|------|----------|----------|----------|
| Trigger | Auto | External | External | Auto |
| Charge measured | No | Yes | Yes | Optional |
| Zero-suppression | - | No | No | Optional |
| Fixed data size | Yes | Yes | Yes | No |




Fig. 21. TRT measurement as a function of the threshold set in one channel of MAROC3 with no charge injected (red) and with ≈ 53 fC injected at 50 kHz (blue).

calibrate it and perform data acquisition. Each of those modes can also have additional configurable options for added flexibility, while keeping the same global goal. The four modes foreseen, which will be described in details in the following subsections, are the trigger rate test (TRT), pedestal measurement (PED), LED light injection (LED), and normal data taking modes. Table 2 shows some characteristics for these four modes.

6.1. Trigger rate test mode

This mode is used to measure the trigger rate per channel. The main goals of this mode are to characterise the electronic noise, define the trigger threshold and search for light leaks in the TT modules. In the TRT mode, the acquisition is set in auto-trigger mode without charge acquisition, only the number of triggers in a given time window (typically 1 s) for each channel is returned. It is also foreseen that the trigger rate of all channels will be monitored during data taking automatically, in addition to this acquisition mode. The main advantage of having the possibility to request this data is to reduce the latency in this measurement for direct usage during the above-mentioned calibration and search for light leaks.

Fig. 21 shows an example of the determination of the threshold and electronic noise for one particular channel of the FEB using multiple acquisitions using the TRT mode with a varying threshold set in the FEB. For the determination of the electronic noise no charge is injected, while for the determination of the threshold a known charge (corresponding to $1/3$ p.e.) is injected in the tested channel. Both measurements have been done, using a test setup, for all channels of all FEB cards to validate the FEB production. In all accepted cards, a clear distinction between the electronic noise (red) and the $1/3$ p.e. injected charge (blue) is observed, with a change of the 50% threshold from 357 DAC units to 420 DAC units in Fig. 21. The exact thresholds are not at the same DAC values for each FEB and have to be set individually from the values obtained during testing. For all accepted FEB, the standard deviation of threshold values among the 64 channels in the board is in average 6.5 DAC units, and it is always below 10 DAC units.




Fig. 22. Trigger frequency versus channel for one sensor before (red) and after (blue) fixing noisy channels for one of the cards of the TT prototype.




Fig. 23. Mean (left) and standard deviation (right) of pedestal data for one MA-PMT used in the TT prototype. Wilkinson charge readout used.

In addition to determining the desired operational threshold, the TRT mode can be used to check noisy channels during normal operation. The presence of noisy channels can be induced, for example, by light leaks inside the detector. Fig. 22 shows an example of such a measurement done on the TT prototype (described in Section 7) before and after fixing a grounding issue due to an improper installation of the ROB.

6.2. Pedestal mode

This mode is used to determine the measured charge in the absence of a signal. This is achieved by measuring the charge with a fixed trigger frequency rather than using the MA-PMT signal to trigger the detection. It is possible that some of the fixed triggers will coincide in time with regular triggers from the MA-PMT by chance. These rare coincidences will produce events with charge away from the pedestal peak and can easily be removed from the pedestal distribution when the pedestal run is being analysed. Fig. 23 presents the pedestal values of the 64 channels of one sensor (left) and their standard deviation values (right). The mean of the pedestal standard deviation, using the 8-bit Wilkinson charge readout mode, is of the order of 0.3 ADC counts, representing about 0.04 p.e.

6.3. LED light injection mode

This mode works similarly to the Pedestal mode described before, however the measurement happens at about the same time of the light injection. As in the Pedestal mode, in the LED mode data is taken with a fixed trigger frequency, with which the LED is pulsed ignoring the MA-PMT trigger.

The light injected in the system uses the same system described in [5]. In this system, light is injected at each end-cap into the WLS




Fig. 24. 3D view of the central part of an end-cap.




Fig. 25. Charge distribution using the light injection system.

fibres just in front of the fibre-MA-PMT opto-coupler with the help of two blue LEDs, straight PMMA light guides and a white painted diffusive box as shown in Fig. 24.

The LEDs are pulsed by a signal from the ROB, which is equipped with a driver designed for this application, as described previously. After a certain configurable delay from the LED pulse, to account for the time required to flash the LED, an external trigger is generated to perform the charge measurement in the MAROC3. The LED intensity can be configured to produce a signal with a mean intensity around the single p.e. as shown in Fig. 25 with the digitised charge readout for the 8-bit Wilkinson mode. In this figure a fit to the single p.e. distribution using the function described in [17] is also drawn. In this distribution the peak related to the pedestal is still clearly visible, with an additional set of distributions corresponding to the cases where one to five p.e. have been detected by that MA-PMT channel. In this case, the measured mean intensity of the light (μ) corresponds to 1.92 ± 0.02 p.e. and the 1 p.e. measured charge ($Q1$) is 11.29 ± 0.09 ADC units. This measured charge is used to calibrate the conversion between ADC units and charge as the PMT gain had been set to 10^6 .

While the system is designed to produce an approximately uniform illumination of all fibres, since the illumination is done from the sides, the fibres at the edge of the opto-coupler receive more light and shadow the more central fibres. Due to these geometrical effects, the illumination in the different fibres might vary by a factor of 3, as shown in Fig. 26. For an optimal performance of the fit described before, it is suitable to tune the light intensity to have the average number of photons observed in the 1–2 p.e. range [18]. In this range, both the pedestal and the single photo-electron peak are clearly visible.

The light injection system can then be used to define the gain equalisation factor to be used in the MAROC3 to equalise the gain in




Fig. 26. Mean light intensity per channel using the light injection system for one MA-PMT. Variations of the mean intensity are due to geometrical effects.




Fig. 27. Gain per channel measured using the light injection system. Red empty (blue full) markers show the gain before (after) the gain equalisation procedure.

all channels of the MA-PMT. The measured gain in each channel before (red) and after (blue) gain equalisation using these gain equalisation factors are shown in Fig. 27. Before equalisation, gains vary by a factor of up to 3, while after this procedure gains vary only by 30%.

This LED mode is foreseen to be used regularly during the whole duration of the experiment, mainly to monitor the MA-PMT gain.

6.4. Normal run mode

This mode is the one used for regular data acquisition. In this case the MA-PMT signal passing the threshold will produce a MAROC3 trigger as described beforehand. Additionally, zero-suppression in the ROB can be enabled in this mode to reduce the data size.

To safeguard against unexpectedly high rates, it is also possible to entirely disable the charge readout, which reduces the detector dead-time and significantly decreases the data size, while keeping the information about the triggered channels. A mixed acquisition option with and without charge readout is being investigated. In this case, a new trigger will start charge readout if the acquisition system is not already performing the charge readout for another trigger. However, in the case where the acquisition system is busy, no charge readout will be performed and only the triggered channels and trigger time will be saved.

The data produced by the TT readout system in the normal mode contains the MA-PMT trigger time-stamp provided by the CB, the trigger register identifying which channels on the MA-PMT triggered, and the measured charges if charge readout is enabled.




Fig. 28. Schematic view of TT prototype.

7. The TT prototype

In order to test the electronics under development, a prototype of the TT was built at IPHC⁹ using the worst modules produced for the OPERA experiment. It is also equipped with the same electronics cards as the TT for testing.

This prototype, shown in Fig. 28, is made of 4 layers. As in the case of the TT, each layer is composed of two planes oriented perpendicularly. Each plane on the prototype was built by cutting a TT module in 3 parts at a distance of about 1.7 m from each end, with the left and right end-caps becoming the x and y oriented planes, respectively, which are used for the x - y coincidences as in the TT. This means that each layer of the prototype was made from a single TT module and has a surface of about 1/16 of the surface of one TT wall. Differently from the TT modules which are read on both sides, the prototype ones are only read on one side.

In each of the 8 modules end-caps was installed one MA-PMT, one FEB, and one ROB. All the 8 ROB are connected to one CB placed in the metal frame supporting the prototype. In the prototype it is also possible, for testing purposes, to read the data directly from the ROB without passing by the CB, by using a different version of the ROB specifically designed for that purpose. Using the prototype, it has already been possible to validate the calibration procedures and test the interface between the various electronics cards using natural signals, albeit with smaller rate than the one expected in JUNO, due to the lack of the high rate of events expected from the rock radioactivity. It is worth noting that most figures in Section 6 were obtained from data of this prototype.

Using the TT prototype several consecutive days of data have been collected. This data makes it possible to check, e.g., the difference in the coincidence time between the two planes of the same layer, as shown in Fig. 29. In this figure the average time difference between the MA-PMT trigger on the left and right cards is shown as a function of the channel triggered in each card. This distribution reflects the different length of the fibres from each interaction point to the MA-PMT for the x and y directions.




Fig. 29. Distribution of the average time difference between signals in left and right cards as a function of the strip which triggered the MA-PMT in each card during 'normal' data taking.




Fig. 30. Distribution of the direction of muons observed in the prototype. The 3D sphere is projected in a plane using the Mollweide projection. Longitude lines are shown every 30°, while latitude lines are shown every 22.5°. The prototype is facing upwards during this data taking. Colour code corresponds to the observed muon rate.

Using the data from the TT prototype it is possible to measure the atmospheric muon flux distribution at IPHC by requiring coincidences between 3 or 4 layers of the prototype, and reconstructing the muon tracks. Fig. 30 shows the reconstructed muon directions obtained using the Mollweide projection. Given the requirement of 3 or 4 layers in coincidence for the reconstruction, the field of view of the telescope is limited to about 70° from the vertical. A motor is added to the prototype rotation axis as shown in Fig. 28 to increase this field of view by rotating the whole device.

In order to have a good estimate of the timing capabilities the prototype has been placed in vertical position (i.e., a 90° rotation with respect to the position shown in Fig. 28) to record horizontal atmospheric muons. In this position, it is expected to record about the same number of muons coming from each side. Fig. 31 presents the time difference between hits recorded on first and last layers of the prototype. The distance between these two layers being of 90 cm it is expected to observe a time difference of the order of 3 ns assuming all muons travel at the speed of light. The two peaks observed on Fig. 31 correspond to muons coming from either side of the detector. They are centred at about ± 3 ns, with 0 corresponding to events being seen at the same time in both layers. The time dispersion is of the order of 2 ns and includes the decay time in the scintillator (2.3 ns) and the WLS

⁹ Institut Pluridisciplinaire Hubert CURIE, Strasbourg, France.




Fig. 31. Time difference between the first and last layers of the prototype from atmospheric muons when the prototype is in vertical position.



Fig. 32. The TT wall, measuring $7.6 \times 7.6 \text{ m}^2$, assembly on its support frame called a “table”.

fibres (7.6 ns), and the time-walk in the electronics threshold. All other delays, as the difference in length of fibres from one channel to another, have been corrected. The time-bin of the clock used for timestamping triggers during this data taking was of 1.04 ns. The slight asymmetry between the two peaks is probably due to the other objects around the prototype.

8. Installation

The 496 TT modules were transported from LNGS to the JUNO site in 2017. According to the schedule, the installation of the TT will begin after the assembly of the central detector, presumably end of 2023.

A dedicated support structure providing the necessary rigidity has been designed (Fig. 4). It includes two main beams of 48 m across the experimental hall, 3 layers of the transverse girders of 20 m between them, a support of the JUNO calibration system, and the support frames of the TT walls.

The walls will be assembled on their support frames, called “tables”. 4 modules will be placed on the table in one direction (x) and other 4 modules in the perpendicular direction (y), as shown in Fig. 32.

All the modules will be equipped with the FEB and ROB and commissioned before installation. Then, each wall will be equipped with




Fig. 33. A scheme of the installation of the TT “tables” in the TT support.

the CB and with the cables to connect all ROB to the CB. Assembled this way, the walls will be commissioned (the counting rate of each MA-PMT should be similar and corresponding to the local atmospheric muons and radioactivity rate) and then moved by a crane to their final place in the detector. The tables are rigid enough to keep the walls flat during the manipulations by the crane and in the detector after the installation where they rest on the transverse girders, as shown in Fig. 33.

In total, 60 walls will be installed in 3 layers separated by 150 cm in height. To avoid inefficiency in the detection of atmospheric muons crossing regions between neighbouring walls, the walls overlap by 150 mm on all sides. For that, the overlapping walls have slightly different height of their position and have to be installed in a proper order. In addition to that, the “tables” of the three walls above the chimney have a different design providing a separation by only 21 cm in their vertical position. During the installation, the position of the walls will be measured by survey instruments with an accuracy of about 1 mm. The relative alignment of the walls can be further verified with the help of atmospheric muons during the data taking.

According to the installation schedule, one wall per day will be assembled and mounted in the detector. This time includes some time for curing the modules of any light leaks that might have appeared due to transportation or storage conditions.

9. Ageing monitoring

The production of the TT strips have been done by AMCRYS-H company [19] during two years from 2004 to 2006. It is expected that some ageing of the plastic scintillator will occur before the TT will be used in JUNO experiment, and will continue during the whole lifetime of the detector.

From August 2006 until April 2013 electronic cards of the OPERA experiment were almost continuously registering atmospheric muons along with muons produced in ν_μ CC interactions during the OPERA beam (CERN Neutrinos to Gran Sasso, CNGS) runs. Over the entire period of the data taking, more than 10^6 straight muon tracks were reconstructed in the Target Tracker detector. Those tracks were used, in particular, to measure the TT module detection efficiency and to monitor the stability of the TT response to minimum ionising particles. The TT modules detection efficiency at 1/3 p.e. threshold was measured to be equal to $(98.5 \pm 0.3)\%$ in “OR” mode (i.e., when a signal above the threshold is required from at least one side of a scintillator strip) [20]. The ageing of the plastic scintillator was estimated using the time evolution of the most probable values of the TT response distributions shown in Fig. 34 obtained for the periods of five years CNGS runs (2008–2012) [21].

The characteristic value of the ageing was found to be $(1.7 \pm 0.2)\%$ light yield loss per year, which was in agreement with indirect estimates of the plastic scintillator manufacturer ($\sim 20\%$ over 11.9 years) [22]. In fact, this evaluation corresponds to the ageing of the entire




Fig. 34. Distribution of the OPERA Target Tracker signal amplitudes, the fitted line is a convoluted Landau and Gaussian distribution. Data used for this distribution was taken during CNGS runs (2008–2012).




Fig. 35. Variation of the most probable value (MPV) of the TT amplitudes with time. This data taking period corresponds to when the TT modules were stored in China prior to their assembly in JUNO. In early 2019 the TT was moved from a warehouse in Zhongshan to the JUNO site. The red line corresponds to the amplitude expected by extrapolating the measured amplitude loss during the CNGS runs.

system also including possible changes of properties of the other TT strips components and materials, like the WLS fibres, glue, etc.

To follow the ageing of the TT sensitive materials after dismantling of the OPERA detector and until its use in JUNO, some modules remained equipped with the old OPERA electronics during their transportation to China and subsequent storage. These modules are used to register atmospheric muons traversing the detector while it remains in storage prior to it being mounted in JUNO. This made it possible to still register atmospheric muons with fewer modules. Although during the period of TT storage, the layout of the containers with the modules have been changed several times, the data collected during each period demonstrated stable average signal amplitude for passing through muons. It is worth noting that while the change in the layout of the containers can create a systematic change in the amplitude of the signal for each period, as it was observed when the containers were moved in early 2019, the comparisons of the measurements within each period allow us to measure the relative performance deterioration. The analysis of the collected data shows no additional deterioration of the performance of the plastic scintillator compared to already mentioned ageing (Fig. 35).

Finally, from their production, about 20 years ago, up to the beginning of the JUNO data taking, about 20% decrease in the number of

p.e. is expected. Considering that 6 p.e. were observed at the middle of each strip (most disfavoured position) after their production, 4.8 p.e. are expected at the beginning of the JUNO experiment. With 1/3 p.e. threshold, the muon detection efficiency remains at comparable level as the initial one. In case more p.e. are needed during the JUNO data taking, there is the possibility to introduce optical grease between the fibre opto-coupler and the MA-PMT glass, something which has not been done up to now because it was not needed. Tests have shown that this will increase by about 15% the number of detected p.e., compensating a large part of the loss due to ageing.

10. Summary and conclusions

The OPERA experiment Target Tracker has been operated from 2006 to 2013 using the CNGS neutrino beam and atmospheric muons. After decommissioning, it has been sent to China to be used as Top Tracker (TT) for the neutrino oscillation JUNO experiment. The available $6.7 \times 6.7 \text{ m}^2$ TT walls will be distributed in 3 layers of a 3×7 horizontal grid above the water Cherenkov detector. The TT will help to study the background created by the passage of atmospheric muons and hence reduce the systematic error coming from ^9Li and ^8He decays mimicking IBD in the JUNO central detector. Despite not covering the entire surface above the central detector, the TT will help the other parts of the detector to tune their algorithms to improve their reconstruction of atmospheric muons and veto the affected regions.

For this utilisation, the TT electronics have been modified to cope with the high background induced by the rock radioactivity. These electronics include for each sensor a Front End Board and a Readout Board. All 16 sensors of a TT wall are connected to a Concentrator Board whose role is to emit a first trigger level by performing x–y coincidences in the same TT wall. The Concentrator Boards communicate with a Global Trigger Board providing a three-layer TT trigger and with the JUNO data acquisition system. The trigger rate of the TT after the Global Trigger Board is expected to be around 2 kHz, pushed down to a few Hz by an offline muon track reconstruction.

In order to validate the whole electronics chain and trigger strategy a prototype has been constructed using the same elements as the Top Tracker.

The TT has been sent to China in 2017 and stored in a hall with a temperature lower than $25 \text{ }^\circ\text{C}$ to keep the plastic scintillator ageing to a reasonable level. This ageing is expected to cause a loss of scintillation light of the order of 1%/year. Waiting for the TT installation in the underground JUNO site, the scintillator ageing is monitored using atmospheric muons. Up to now, the observed ageing is compatible with the expectation.

The installation of the TT is expected to take place at the end of 2023. It is evaluated that this installation will last for less than six months.

Declaration of competing interest

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

Data availability

No data was used for the research described in the article.

Acknowledgements

We are grateful for the ongoing cooperation from the China General Nuclear Power Group. We acknowledge the supported of the Chinese Academy of Sciences, the National Key R&D Program of China, the CAS Center for Excellence in Particle Physics, Wuyi University, and the Tsung-Dao Lee Institute of Shanghai Jiao Tong University in China, the

Institut National de Physique Nucléaire et de Physique de Particules (IN2P3) in France, the Istituto Nazionale di Fisica Nucleare (INFN) in Italy, the Italian-Chinese collaborative research program MAECI-NSFC, the Fond de la Recherche Scientifique (F.R.S-FNRS) and FWO under the “Excellence of Science – EOS” in Belgium, the Conselho Nacional de Desenvolvimento Científico e Tecnológico in Brazil, the Agencia Nacional de Investigacion y Desarrollo in Chile, the Charles University Research Centre and the Ministry of Education, Youth, and Sports in Czech Republic, the Deutsche Forschungsgemeinschaft (DFG), the Helmholtz Association, and the Cluster of Excellence PRISMA+ in Germany, the Joint Institute of Nuclear Research (JINR) and Lomonosov Moscow State University in Russia, the joint Russian Science Foundation (RSF) and National Natural Science Foundation of China (NSFC) research program, the MOST and MOE in Taiwan, the Chulalongkorn University and Suranaree University of Technology in Thailand, and the University of California at Irvine in USA. This work, in particular, was supported by the initiative of excellence IDEX-Unistra (ANR-10-IDEX-0002-02) and EUR-QMat (ANR-17-EURE-0024) from the French national program “investment for the future”, and by the Strategic Priority Research Program of the Chinese Academy of Sciences, Grant No. XDA10010300.

We would like to thank all private companies which have collaborated with our institutes in developing and providing materials. We are particularly grateful to the OPERA Collaboration for the donation of all TT modules to the JUNO Collaboration. Finally, we would like to thank all technicians non-authors of this paper who have helped us all these last years.

References

- [1] Z. Djurcic, et al., JUNO conceptual design report, 2015, [arXiv:1508.07166](https://arxiv.org/abs/1508.07166).
- [2] A. Abusleme, et al., JUNO physics and detector, *Prog. Part. Nucl. Phys.* 123 (2022) 103927, [http://dx.doi.org/10.1016/j.pnpnp.2021.103927](https://doi.org/10.1016/j.pnpnp.2021.103927), [arXiv:2104.02565](https://arxiv.org/abs/2104.02565).
- [3] F. An, et al., Neutrino physics with JUNO, *J. Phys. G* 43 (3) (2016) 030401, [http://dx.doi.org/10.1088/0954-3899/43/3/030401](https://doi.org/10.1088/0954-3899/43/3/030401), [arXiv:1507.05613](https://arxiv.org/abs/1507.05613).
- [4] M. Guler, et al., OPERA: An appearance experiment to search for $\nu/\mu \leftrightarrow \nu/\tau$ oscillations in the CNGS beam, *Exp. Propos.* (2000).
- [5] T. Adam, et al., The OPERA experiment target tracker, *Nucl. Instrum. Methods A* 577 (2007) 523–539, [http://dx.doi.org/10.1016/j.nima.2007.04.147](https://doi.org/10.1016/j.nima.2007.04.147), [arXiv:physics/0701153](https://arxiv.org/abs/physics/0701153).
- [6] D. Malczewski, et al., Gamma background measurements in the gran sasso national laboratory, *J. Radioanal. Nucl. Chem.* 295 (2013) 749–754, [http://dx.doi.org/10.1007/s10967-012-1990-9](https://doi.org/10.1007/s10967-012-1990-9).
- [7] K. Li, et al., GDML based geometry management system for offline software in JUNO, *Nucl. Instrum. Methods A* 908 (2018) 43–48, [http://dx.doi.org/10.1016/j.nima.2018.08.008](https://doi.org/10.1016/j.nima.2018.08.008).
- [8] S. Agostinelli, et al., GEANT4—a simulation toolkit, *Nucl. Instrum. Methods A* 506 (2003) 250–303, [http://dx.doi.org/10.1016/S0168-9002\(03\)01368-8](https://doi.org/10.1016/S0168-9002(03)01368-8).
- [9] S. Blin, et al., MAROC, a generic photomultiplier readout chip, in: *IEEE Nucl. Sci. Symp. Conf. Rec.*, Vol. 2010, 2010, pp. 1690–1693, [http://dx.doi.org/10.1109/NSSMIC.2010.5874062](https://doi.org/10.1109/NSSMIC.2010.5874062).
- [10] ANALOG DEVICES, AD9629 datasheet, 2022, URL <https://www.analog.com/en/products/ad9629.html#product-overview> (accessed on 11/01/2022).
- [11] CAEN, A7511 webpage, 2022, URL <https://www.caen.it/products/a7511/> (accessed on 11/01/2022).
- [12] M. Lipiński, et al., White rabbit: a PTP application for robust sub-nanosecond synchronization, in: *2011 IEEE International Symposium on Precision Clock Synchronization for Measurement, Control and Communication*, 2011, pp. 25–30, [http://dx.doi.org/10.1109/ISPCS.2011.6070148](https://doi.org/10.1109/ISPCS.2011.6070148).
- [13] M. Lipinski, et al., White-rabbit node, 2022, URL <https://ohwr.org/project/white-rabbit/wikis/Node> (accessed on 04/07/2022).
- [14] AMD XILINX, Zynq UltraScale+ mp soc, 2022, URL <https://www.xilinx.com/products/silicon-devices/soc/zynq-ultrascale-mpsoc.html> (accessed on 04/07/2022).
- [15] Image Matters, Origami module B20, 2018, URL https://solutions.inrevium.com/products/pdf/ORIGAMI_datasheet_A4_2018Mar.pdf (accessed on 29/07/2021).
- [16] AMD XILINX, UltraScale architecture and product data sheet: Overview, 2022, URL https://www.xilinx.com/content/dam/xilinx/support/documentation/data_sheets/ds890-ultrascale-overview.pdf (accessed on 04/07/2022).
- [17] E.H. Bellamy, et al., Absolute calibration and monitoring of a spectrometric channel using a photomultiplier, *Nucl. Instrum. Methods A* 339 (1994) 468–476, [http://dx.doi.org/10.1016/0168-9002\(94\)90183-X](https://doi.org/10.1016/0168-9002(94)90183-X).
- [18] N. Anfimov, et al., Optimization of the light intensity for photodetector calibration, *Nucl. Instrum. Methods A* 939 (2019) 61–65, [http://dx.doi.org/10.1016/j.nima.2019.05.070](https://doi.org/10.1016/j.nima.2019.05.070), [arXiv:1802.05437](https://arxiv.org/abs/1802.05437).
- [19] AMCRYS-H, Scintillation materials, detectors and electronics, 2022, URL <https://www.amcrys.com> (accessed on 11/01/2022).
- [20] S. Dmitrievsky, Поиск нейтринных взаимодействий и исследование свойств нейтрино с помощью электронных детекторов в эксперименте OPERA (Search for Neutrino Interactions and Investigation of Neutrino Properties with the Help of Electronics at the OPERA Experiment) (Ph.D. thesis), 2015, URL <http://operaweb.lngs.infn.it/Opera/ptb/theses/theses/Dmitrievsky-Sergey-PhD-thesis.pdf> (accessed on 22/11/2022).
- [21] CERN, CNGS webpage, 2022, URL <http://proj-cngs.web.cern.ch/proj-cngs> (accessed on 21/11/2022).
- [22] B.V. Grinev, V.G. Senchyshyn, Пластмассовые сцинтилляторы Plastic Scintillators, Acta, Kharkiv, ISBN: 966-7021-67-X, 2003.