


CLAS12-RICH Status-Report

September 5th 2014





AEROGEL:

Investigating X-ray camera analysis

Dark+dry box ready

Optimizing box humidity control

Upgrading automatize test-bench motor control

Rotating bench delivered

Next step: - Systematic measurements



MIRRORS:

Measures of surface roughness done in Frascati to be validated by Media-Lario

Preliminary results from D0 measurement done in Frascati:

- strong aberration at the edges of the CMA samples
- no measurable D0 for Riba samples

Surface measurements of Riba samples done at Ferrara

Next step: - Coating at ZAOT and SESO-Thales

- 2nd CMA demo on Marcon mandrel



PHOTON-DETECTORS:

Detailed analysis of H12700 ongoing at Jlab
~ 25% better photon yield vs H8500
worse SPE signal vs H8500
(better QE, worse collection efficiency ?)



ELECTRONICS:

Delivery of prototype boards expected soon
Meeting next week to review status and plan electronic tests



SIMULATIONS:


RICH implementation into GEMC 2.0 ongoing




WEB PAGES:

New link on the CLAS12 wiki (reorganized RICH Review page)

Aerogel: X-ray Camera Scan




Aerogel: X-ray Camera Scan




Aerogel: Black+Dry Box




Aerogel: Humidity Control




Aerogel: Humidity Control



Humidity Control



Aerogel: Dry Cabinet



3% stable humidity during winter

Possible improvements: external dehumidifier + internal fan